



NEW DESIGN Tc-99m GENERATOR



DO NOT 



Do not TOUCH and REMOVE needle protector (Plunger).

The needle protector (Plunger) should only come into contact with the new elution shield.

Do not remove elution shield at an angle to the side. The elution shield should be removed upright.

Do not open it by pressing a single button. Push the opening buttons simultaneously and removes the cap from outer container easily. User closes the cap with one move and secures it to the container by hearing the click sound.

DO 

